



RADIOLOGY EQUIPMENT - NW EXPANSION

VENDOR QUESTIONS

Q1: Is a CAD license needed for the GE Pristina unit?

A1: No, we do not need a CAD license.

Q2: Please clarify the quantity of Logiq E10 ultrasound units required; page 3 list quantity of two(2) but exhibit A shows quantity of (1)?

A2: There was an error on page 3. We only need 1 ultrasound unit. Please see revised list as a 3rd piece of equipment has been added.

Q3: Which probes are required for the ultrasound unit(s)?

A3: L2-9-DXD-Clear, ML4-20, C1-6-D, and IC5-9-D.

Q4: Are any additional packages required for the ultrasound unit(s), i.e. Stress Echo, ShearWave, Cardiac, etc.?

A4: Auto abdominal package.

Q5: Please clarify the specifics of the "Design Services" requirement; does this require services outside of the standard room drawings and shielding plans?

A5: This would include drawings, shielding plans and coordinating/verifying needed power and other requirements with the architects and contractors.

Q6: Once decisions are made, when do you intend to actually process the order(s) as your estimated delivery date will most like be in late 2025.

A6: We need information on lead times to understand when to place the order. We anticipate wanting equipment install around late October early November of 2025.